Passive 1x/10x Voltage Probe

P2220 P2221 Datasheet



Tektronix passive probe solutions provide high reliability and testing accuracy for basic oscilloscope measurements. The P2220 and P2221 are compact passive probes with selectable 1X/10X attenuation designed for use with selected Tektronix oscilloscopes.

Key features

- 1x/10x attenuation
- 200 MHz bandwidth
- 1.5 m length
- UL3111-1, CSA1010.1, EN61010-1, EN61010-2-031, and IEC61010-2-031

Applications

- Relative low-frequency measurements
- Low-frequency computer and telecom measurements
- Power supplies
- Low-frequency amplifiers

P2220 P2221 passive voltage probes

The P2220 and P2221, 200 MHz passive voltage probes permit selection of 1x or 10x attenuation, using a switch located on the probe head.

The P2220 probe is compatible with TDS200, TDS1000, TDS1000B, TDS2000, TDS2000B, and TPS2000 Series oscilloscopes. The P2220 probe also provides floating measurement capability up to 30 V_{RMS} with TPS2000 Series oscilloscopes only. (Note: Do not float with TDS2000, TDS2000B, TDS1000, TDS1000B, or TDS200 Series oscilloscopes.)

The P2221 is compatible with MSO2000 and DPO2000 Series oscilloscopes.

Specifications

Specifications apply to both probes unless indicated otherwise.

		10X	1x
Bandwidth		DC to 200 MHz	DC to 6 MHz
Probe length		1.5 m	1.5 m
Attentuation ratio		10:1	1:1
Compensation range			All oscilloscopes with 1 MΩ input
	P2220	15 to 25 pF	
	P2221	10 to 25 pF	
Input capacitance		17.0 pF	110 pF
Input resistance		10 ΜΩ	1 ΜΩ



Datasheet

	10X	1x
Maximum input voltage	300 V Cat II	150 V Cat II
Maximum voltage between reference lead and earth ground	30 V _{RMS} ¹	30 V _{RMS} ¹

Ordering Information

P2220 Passive Probe, 1x/10x, 200 MHz Passive Voltage Probe

Includes retractable hook tip (013-0107-xx), 6-inch ground lead (196-3466-xx), adjustment tool (003-1433-xx), marker band set

(016-1315-xx), instruction manual (071-1464-xx). 2

P2221 Passive Probe, 1x/10x, 200 MHz Passive Voltage Probe for MSO/DPO2000 Series oscilloscopes

Includes retractable hook tip (013-0107-xx), 6-inch ground lead (196-3466-xx), adjustment tool (003-1433-xx), marker band set

(016-1315-xx), instruction manual (071-2335-xx).2

 ϵ

ASEAN / Australasia (65) 6356 3900
Belgium 00800 2255 4835*
Central East Europe and the Baltics +41 52 675 3777
Finland +41 52 675 3777
Hong Kong 400 820 5835
Japan 81 (3) 6714 3010
Middle East, Asia, and North Africa +41 52 675 3777
People's Republic of China 400 820 5835
Republic of Korea 001 800 8255 2835
Spain 00800 2255 4835*
Taiwan 886 (2) 2722 9622

Taiwan 886 (2) 2722 9622

* European toll-free number. If not accessible, call: +41 52 675 3777

Brazil +55 (11) 3759 7627 Central Europe & Greece +41 52 675 3777 France 00800 2255 4835* India 000 800 650 1835 Luxembourg +41 52 675 3777

Austria 00800 2255 4835*

The Netherlands 00800 2255 4835* Poland +41 52 675 3777 Russia & CIS +7 (495) 6647564 Sweden 00800 2255 4835*

United Kingdom & Ireland 00800 2255 4835*

Balkans, Israel, South Africa and other ISE Countries +41 52 675 3777

Canada 1 800 833 9200 Denmark +45 80 88 1401 Germany 00800 2255 4835* Italy 00800 2255 4835*

Mexico, Central/South America & Caribbean 52 (55) 56 04 50 90

Norway 800 16098 Portugal 80 08 12370 South Africa +41 52 675 3777 Switzerland 00800 2255 4835* USA 1 800 833 9200

Updated 10 April 2013

For Further Information. Tektronix maintains a comprehensive, constantly expanding collection of application notes, technical briefs and other resources to help engineers working on the cutting edge of technology. Please visit www.tektronix.com.

Copyright © Tektronix, Inc. All rights reserved. Tektronix products are covered by U.S. and foreign patents, issued and pending. Information in this publication supersedes that in all previously published material. Specification and price change privileges reserved. TEKTRONIX and TEK are registered trademarks of Tektronix, Inc. All other trade names referenced are the service marks, trademarks, or registered trademarks of their respective companies.

19 Apr 2013

51W-17984-2

www.tektronix.com



¹ The float voltage must be subtracted from the tip to earth ground voltage. For example, if the reference lead is floated to 30 V RMS while the probe is in the 10x position, the tip voltage to the reference lead is limited to 270 V PMS.

² The accessory quantities may differ when multiple probes are shipped with an oscilloscope/probe system.